



### **Roydent Engine Driven Instruments Sterilization Chart**

Gates Glidden, Peeso Reamers, Engine Hedstroms, Ending Reamers

Type	Comments
Hot Air +C° max, 180°	Yes
Autoclave Water Steam	Yes
Autoclave Chemical Vapor	Yes
Cold Sterile, Disinfectant Bath	Possible, a variable reaction (from zero to strong) may take place, depending upon the concentration of the utilized product. The result is a fading of the color markings.