



Roydent Hand Instruments Sterilization Chart
Hand Files, Barbed Broaches

Type	Comments
Hot Air +C° max, 180°	Yes
Autoclave Water Steam	Yes
Autoclave Chemical Vapor	Yes
Cold Sterile, Disinfectant Bath	Possible, see observations
Flame Heat	Stainless steel shaft only, may curl in direct heat
Observations	A variable reaction (from zero to strong) may take place, depending upon the concentration of the product utilized. The result is fading of any color markings.